
NI-9860

Specifications

Ihr NI-Partner: 

AMC – Analytik & Messtechnik GmbH Chemnitz

Heinrich-Lorenz-Str. 55 Tel.: +49/371/38388-0
09120 Chemnitz Fax: +49/371/38388-99
E-Mail: info@amc-systeme.de Web: www.amc-systeme.de



Integration
Partner

SYSTEM INTEGRATOR



Contents

NI-9860 Specifications..... 3

NI-9860 Specifications

Module Version

These specifications apply to NI-9860 modules with part number 157128x-04L (where x represents any letter).

Definitions

Warranted specifications describe the performance of a model under stated operating conditions and are covered by the model warranty.

Characteristics describe values that are relevant to the use of the model under stated operating conditions but are not covered by the model warranty.

- **Typical** specifications describe the performance met by a majority of models.
- **Nominal** specifications describe an attribute that is based on design, conformance testing, or supplemental testing.

Specifications are **Typical** unless otherwise noted.

Conditions

Specifications are valid for the range -40 °C to 55 °C unless otherwise noted.

Related information:

- [Software Support for CompactRIO, CompactDAQ, Single-Board RIO, R Series, and EtherCAT](#)

NI-XNET Host Port

| | |
|--|--|
| Connector type | NI-XNET hardware selectable interface port |
| Port supported transceiver cable types | NI-XNET transceiver cables (CAN/LIN) |



Note For more information about transceiver cables, refer to the transceiver cable operating instructions.

Safety Voltages

Connect only voltages that are within the following limits:

| | |
|-----------------------|--|
| Vsup-to-COM | 9 V DC to 30 V DC, 1.5 W maximum, Measurement Category I |
| Chassis ground-to-COM | 30 V DC maximum, Measurement Category I |

Measurement Category I

Warning Do not connect the product to signals or use for measurements within Measurement Categories II, III, or IV, or for measurements on MAINS circuits or on circuits derived from Overvoltage Category II, III, or IV which may have transient overvoltages above what the product can withstand. The product must not be connected to circuits that have a maximum voltage above the continuous working voltage, relative to earth or to other channels, or this could damage and defeat the insulation. The product can only withstand transients up to the transient overvoltage rating without breakdown or damage to the insulation. An analysis of the working voltages, loop impedances, temporary overvoltages, and transient overvoltages in the system must be conducted prior to making measurements.

Mise en garde Ne pas connecter le produit à des signaux dans les catégories de mesure II, III ou IV et ne pas l'utiliser pour des mesures dans ces catégories, ou des mesures sur secteur ou sur des circuits dérivés de surtensions de catégorie II, III ou IV pouvant présenter des surtensions transitoires supérieures à ce que le produit peut supporter. Le produit ne doit pas être raccordé à des circuits ayant une tension maximale supérieure à la tension de fonctionnement continu, par rapport à la terre ou à d'autres voies, sous peine d'endommager et de compromettre l'isolation. Le produit peut tomber en panne et son isolation risque d'être endommagée si les tensions transitoires dépassent la surtension transitoire nominale. Une analyse des tensions de fonctionnement, des impédances de boucle, des surtensions temporaires et des surtensions transitoires dans le système doit être effectuée avant de procéder à des mesures.

Measurement Category I is for measurements performed on circuits not directly connected to the electrical distribution system referred to as **MAINS** voltage. MAINS is a hazardous live electrical supply system that powers equipment. This category is for measurements of voltages from specially protected secondary circuits. Such voltage measurements include signal levels, special equipment, limited-energy parts of equipment, circuits powered by regulated low-voltage sources, and electronics.



Note Measurement Categories CAT I and CAT O are equivalent. These test and measurement circuits are for other circuits not intended for direct connection to the MAINS building installations of Measurement Categories CAT II, CAT III, or CAT IV.

Environmental Characteristics

| Temperature | |
|-------------|-----------------|
| Operating | -40 °C to 55 °C |
| | |

| | |
|--|--|
| Storage | -40 °C to 85 °C |
| Humidity | |
| Operating | 10% RH to 90% RH, noncondensing |
| Storage | 5% RH to 95% RH, noncondensing |
| Ingress protection | |
| NI-9860 | IP30 |
| NI-9860 with power and transceiver cables attached | IP40 |
| Pollution Degree | 2 |
| Maximum altitude | 5,000 m |
| Shock and Vibration | |
| Operating vibration | |
| Random | 5 g RMS, 10 Hz to 500 Hz |
| Sinusoidal | 5 g, 10 Hz to 500 Hz |
| Operating shock | 30 g, 11 ms half sine; 50 g, 3 ms half sine; 18 shocks at 6 orientations |

To meet these shock and vibration specifications, you must panel mount the system.

Power Requirements

| |
|---------------------------------------|
| Power consumption from chassis |
|---------------------------------------|

| | |
|---|---|
| Active mode | 1 W maximum |
| Sleep mode | 5 mW maximum |
| External Power Supply¹ (Vsup) | |
| Voltage input range | 9 V DC to 30 V DC (measured at the NI-9860 power connector) |
| Power consumption | 1.5 W maximum (active mode) |
| Total thermal dissipation on module | 1.2 W maximum (active mode) |

Physical Characteristics

| | |
|------------|--|
| Dimensions | Visit ni.com/dimensions and search by module number. |
| Weight | Approx. 144 g (5.0 oz) |

¹ Required to power NI-XNET Transceiver Cables.